

Clarification to BID doc of Procurement of Medical and Diagnostic equipment and Medical Furniture for "Vanadzor" MC ICB No: CR4/ICB/B-G/013-16

1. Inquiry, regarding item 10 of Lot 1:

As mentioned in LOT-1 "Supply and Installation of Diagnostic and Imaging equipment for "Vanadzor MC" CJSC, item 10 - Floor-mounted digital angiography system for interventional cardiology technical specifications the following features are marked as "possible option".

1. The imaging chain allow advanced 2D and 3D applications (possible as option)

Please confirm that below listed features are also optional, otherwise we need the feature above to be required, or please reject that "The imaging chain allow advanced 2D and 3D applications" is optional feature in requirements.

- Soft tissue imaging with fast reconstruction algorithms
- Dynamic data update during movements of the C-arm, table, zoom, SID and patient movements compensation.
- 3D Acquisition: allows native or subtracted 3D for acquisition of 3D high-contrast images.
- Native 3D reconstruction based on digital rotational angiography with angle triggering.
- High-end post processing workstation, for real-time 3D reconstruction and 3D viewing

2. Respond:

Please kindly be informed that the suggested equipment (item 10 under Lot 1) should have a possibility of 2D and 3D applications, which can be added and used by beneficiary in the future.

Meanwhile, all other mentioned functions are mandatory.