

**ARMENIA**  
**DISEASE PREVENTION AND CONTROL PROJECT**

---

**Credit No. 5222-AM**

**Amendment #1**

**To the Bidding Document for Procurement of:  
Medical, diagnostic equipment, Medical supplies and furniture for “Vanadzor  
MC” CJSC, comprising of 4 lots:**

**ICB No: CR4/ICB/B-G/013-16**

Dear Sir/Madam,

We would like to inform you that we request an amendment to the bidding documents ref: CR4/ICB/B-G/013-16:

Please take the following into consideration in your bid preparation process:

- In the Section VII. Schedule of Requirements- List of goods and delivery schedule- Lot 3. Surgery and Hospital equipment, words “Syringe pump” (of the Line 9) should be read as “Suction pump”.
- In the Section VII. Schedule of Requirements- Technical Specifications- Lot 3. Surgery and Hospital equipment, words “Syringe pump” (of the Line 9) should be read as “Suction pump”. Also, the technical specification should be replaced with the following:

Suction Unit for Major Surgery Procedures.
Mains-powered , mobile on 4 antistatic castors, ABS Casing and 2 graduated Canisters of <b>not less than 2,000ml each</b> made of polycarbonate <b>or equivalent material</b> autoclave able at 121°C and disposable suction bags
Require no maintenance nor lubrication Oil-free Piston Motor, maximum suction 600...700mm Hg Flow rate <b>not less than 45 l/min</b>
Main Switch with Pilot Lamp (fuses)
Equipped with a protective thermal cut-out relay
Equipped with motor-protection cap that totally prevents aspirated liquids or secretions from reaching and damaging the vacuum pump
Suction command with continuous adjustment
Vacuometer: <b>not less than 2x2.000ml</b>
Canisters with airproof screwing-cap with independent
Overflow devices. Fast Connectors and silicone Tubing
PowerSupply: 220V/50Hz.
Sound level: <b>not more than 55dBA</b>
<b>Accessories:</b>
SiliconeTubing, sterilizable
TransparentCannulaHolder, Sterilizable
Anti-BacterialFilters (4)
Set of 4 canulaes with Holder
<b>Standards (Documentary evidence (copy of certificates):</b>
ISO13485 or equivalent
CE Mark (Directive 93/42/EEC) or equivalent

Best Regards,

Nelson Zuloian, HPIU Director

Hovakim Podosyan, HPIU Procurement Specialist