

**Clarification to BID doc of Procurement of Medical and Diagnostic equipment and Medical Furniture
for "Vanadzor" MC , ICB No: CR4/ICB/B-G/013-16**

Crt. No.	Question	Response
1	<p>Lot 2 – item 7 - “Water still”</p> <p>a) Regarding the requirement about overall dimensions of 400x225x545 mm, please accept a deviation of $\pm 10\%$.</p> <p>b) Considering the small overall dimensions, the requested capacity can't be fulfilled in the same time because a water still with a capacity not less than 10 L should have dimension with 60% bigger than requested. Please accept a capacity of 4 L.</p>	<p>a, b) Because of problems with place of installation maximum acceptable dimensions deviation is $\pm 10\%$, meanwhile the capacity should be not less than 10 L.</p>
2	<p>Lot 3 – item 26 “Laparoscopy set”</p> <p>Requirement: 11. Should have bottle for suction, for use with endoscopic aspiration-irrigation pumps 11.1 Volume, l: not less than 5 11.2 Should have construction material – glass</p> <p>Request: Please advise if accepted: - Volume: not less than 3 l - Construction material: hard plastic</p> <p>Requirement: 13. Should have stand for suction bottles 13.1 Should have used with glass aspiration bottles</p> <p>Request: Please advise if accepted: - used with hard plastic aspiration bottles</p> <p>Requirement: 15. Should have foot pedal for irrigation pumps</p> <p>Request: Please advise if accepted: - without foot pedal, with easy controls from irrigation cannulas</p> <p>Requirement: 31. Should have rigid core - lens optics 31.4 Working length, cm: 31</p> <p>Request: Please advise if accepted: - working length: deviation of max. $\pm 10\%$ of 31 cm</p> <p>Requirement: 39-61 Length, cm: 36</p> <p>Request: Please advise if accepted: - Length: deviation of max. $\pm 10\%$</p>	<p>11. 11.1 Volume, should be not less than 5L 11.2 Construction material should be a glass for easier using, cleaning and sterilizing. 13. Same as (11.2 point) 15. Following a practice adopted in our hospitals controls from irrigation cannulas is not acceptable, because of further problems with using. 31. Following a practice adopted in our hospitals deviation in working length is not acceptable. 39-61. Same as (31 point)</p>